## ISLAND SCIENTIFIC LTD Reply to: Unit 9 Ventnor Industrial Estate Station Road



Vacuum Equipment, Used and Reconditioned for Technology, Industry, Research and Education

*Reply to:* Unit 9 Ventnor Industrial Estate Station Road Ventnor *for Technology,* Isle of Wight PO38 1DX England Tel No: (0) 1983 855822 Fax No: (0) 1983 855822 Fax No: (0) 1983 852146 E-Mail Address: sales@island-scientific.co.uk Web Site: http://www.island-scientific.co.uk



## Make: Edwards Model: SP40K

Type: Hand Valve Connection size: KF40 Leak Tightness: 10-6 mbar litres per second Pressure Rating: 10 bar Length: 85 mm Height: 158 mm Weight: 2.3 kg Material: Ali Features: Operation: Manual