

ISLAND SCIENTIFIC LTD

Reply to:
Unit 9
Ventnor Industrial Estate
Station Road
Ventnor
Isle of Wight
PO38 1DX
England
Tel No: (0) 1983 855822
Fax No: (0) 1983 852146
E-Mail Address: sales@island-scientific.co.uk
Web Site: <http://www.island-scientific.co.uk>



*Vacuum Equipment, Used and Reconditioned for Technology,
Industry, Research and Education*



Make: Tylan

Model: FC260

Type: Mass Flow Controller

Length: 125.5 mm

Height: 136 mm

Width: 25.4 mm

Weight: 500 gm

Leak Integrity: 1 x 10⁻⁹ atm cc/sec He

Temperature Range: 5 - 43 °C

Gas Pressure: 2.08 bar (optimal), 10.35 bar (max)

Features: