ISLAND SCIENTIFIC LTD Reply to: Unit 9 Ventnor Industrial Estate Station Road



Vacuum Equipment, Used and Reconditioned for Technology, Industry, Research and Education

Reply to: Unit 9 Ventnor Industrial Estate Station Road Ventnor *for Technology,* Isle of Wight PO38 1DX England Tel No: (0) 1983 855822 Fax No: (0) 1983 855822 Fax No: (0) 1983 852146 E-Mail Address: sales@island-scientific.co.uk Web Site: http://www.island-scientific.co.uk



Make: Edwards Model: PV40PKA

Type: Pipe Line Valve Pressure Range: 1x10-9 to 2100mbar Leak Rate: 10-9 mbar litres per second Operating Temperature: 5 - 100°C Connection size: 40 mm, KF40 Length: 85 mm Width: 76.2 mms Height: 183 mm Material: HE30TF Aluminium Features: 24V Indicator Switches Operation: Pneumatic