

ISLAND SCIENTIFIC LTD

Reply to:

Unit 9
Ventnor Industrial Estate
Station Road
Ventnor
Isle of Wight
PO38 1DX
England

Tel No: (0) 1983 855822

Fax No: (0) 1983 852146

E-Mail Address: sales@island-scientific.co.uk

Web Site: <http://www.island-scientific.co.uk>



*Vacuum Equipment, Used and Reconditioned for Technology,
Industry, Research and Education*



Make: Edwards

Model: PV10PKA

Type: Pipe Line Valve

Pressure Range: 1x10⁻⁹ to 2100mbar

Leak Rate: 10⁻⁹ mbar litres per second

Operating Temperature: 5 - 100°C

Connection size: 10 mm, KF10

Length: 55 mm

Width: 38 mm

Height: 139.5 mm

Material: HE30TF Aluminium

Features:

Operation: Pneumatic