
## ISLAND SCIENTIFIC LTD



Vacuum Equipment, Used and Reconditioned for Technology, Industry, Research and Education

Reply to: Unit 9 Ventnor Industrial Estate Station Road Ventnor Isle of Wight PO38 1DX England Tel No: (0) 1983 855822 Fax No: (0) 1983 855822 Fax No: (0) 1983 855146 E-Mail Address: sales@island-scientific.co.uk Web Site: http://www.island-scientific.co.uk



## Make: Edwards Model: APGL

Type: Pirani Range: 1000 to 10-4 m/bar Connection flange: KF10 Pressure Range: 1x10-4 to 1000 mbar Max Over Pressure: 6 Bar absolute Power Supply: 20 – 36Vdc Power Consumption: 1.5W Output Signal: 2-10Vdc Adjustments: Set Vacuum and Set Atmosphere Temperature Range: 5-60°C (Operating) 0-70° (Storage) Height: 120 mm Width: 56 mm Depth: 34 mm Weight: 0.12 kg Suitable gauge controller: Active Display model supplied separately Features: Integral adjustable set points for process control.